

CALIBRATION LABORATORIES

NVLAP LAB CODE 600160-0


SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

<p><b>HORIBA STEC, Co., Ltd.</b>  <b>Kyoto Fukuchiyama Technology Center</b>                  11-1 Miwa-Cho Miwa                  Fukuchiyama-city, Kyoto prefecture                  JAPAN 620-1445                  Yasuhiro Isobe                  Phone: 81 75 693 2322                  E-mail: yasuhiko.isobe@horiba.com                  URL: <a href="http://www.horiba.com/semiconductor">http://www.horiba.com/semiconductor</a></p>	<p><b>Fields of Calibration</b>                  Mechanical</p>
--	---

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3,5</sup>	Remarks
<b>MECHANICAL</b>			
<b>FLOW RATE (20/M05)</b>			
Gas Mass Flow $q_m$ <sup>Note 7</sup>	$0.006\ 249\ \text{g/min} \leq q_m < 0.012\ 498\ \text{g/min}$ <sup>Note 8, Note 11</sup>	0.12 %	Dynamic Gravimetric Method
	$0.012\ 498\ \text{g/min} \leq q_m \leq 37.495\ \text{g/min}$ <sup>Note 9, Note 11</sup>	0.08 %	
	$37.495\ \text{g/min} < q_m \leq 64.587\ \text{g/min}$ <sup>Note 10, Note 12</sup>	0.20 %	
	$0.006\ 249\ \text{g/min} \leq q_m < 0.012\ 498\ \text{g/min}$ <sup>Note 8</sup>	0.19 %	Calibration against Working Standard
	$0.012\ 498\ \text{g/min} \leq q_m \leq 37.495\ \text{g/min}$ <sup>Note 9</sup>	0.12 %	
	$37.495\ \text{g/min} < q_m \leq 64.587\ \text{g/min}$ <sup>Note 10</sup>	0.30 %	
<b>END</b>			

2024-03-19 through 2025-03-31  
Effective dates

  
For the National Voluntary Laboratory Accreditation Program

CALIBRATION LABORATORIES

NVLAP LAB CODE 600160-0

Notes

**Note 1:** A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

**Note 2:** Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

**Note 3:** The uncertainty associated with a measurement in a CMC is an expanded uncertainty with a level of confidence of approximately 95 %, typically using a coverage factor of  $k = 2$ . However, laboratories may report a coverage factor different than  $k = 2$  to achieve the 95 % level of confidence. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

**Note 3a:** The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

**Note 3b:** As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

**Note 3c:** As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.5 of NIST Handbook 150, Procedures and General Requirements.

**Note 4:** Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

**Note 5:** Values listed with percent (%) are percent of reading or generated value unless otherwise noted.

**Note 6:** NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories

**Note 7:** The capabilities listed are available for the following gases: N<sub>2</sub>, Ar, O<sub>2</sub>, CF<sub>4</sub>, SF<sub>6</sub>, Air, CO<sub>2</sub>, N<sub>2</sub>O. The laboratory may be able to calibrate other inert gases after technical consideration.

2024-03-19 through 2025-03-31

Effective dates



For the National Voluntary Laboratory Accreditation Program

**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 600160-0**

**Notes**

**Note 8:** For example,  $5 \text{ sccm} \leq q_m < 10 \text{ sccm}$  for nitrogen gas ( $\text{N}_2$ ), where:  
sccm is standard  $\text{cm}^3/\text{min}$  for standard conditions of  $0^\circ\text{C}$  and  $101.325 \text{ kPa}$   
slm is standard  $\text{l}/\text{min}$  ( $\text{dm}^3/\text{min}$ ) for standard conditions of  $0^\circ\text{C}$  and  $101.325 \text{ kPa}$

**Note 9:** For example,  $10 \text{ sccm} \leq q_m \leq 30 \text{ slm}$  for nitrogen gas ( $\text{N}_2$ )

**Note 10:** For example,  $30 \text{ slm} < q_m \leq 51.676 \text{ slm}$  for nitrogen gas ( $\text{N}_2$ )

**Note 11:** Supply pressure to Meter under test:  $\leq 650 \text{ kPa}$  absolute, Temperature:  $25.0 \pm 0.5^\circ\text{C}$

**Note 12:** Supply pressure to Meter under test:  $\leq 450 \text{ kPa}$  absolute, Temperature:  $25.0 \pm 0.5^\circ\text{C}$ .

2024-03-19 through 2025-03-31

*Effective dates*

A handwritten signature in blue ink that reads 'Jana S. Laman'.

*For the National Voluntary Laboratory Accreditation Program*