

A. Bowyer, S. Kitchen and R. Maclean

Department of Coagulation, Sheffield Haemophilia and Thrombosis Centre, UK

INTRODUCTION

- The APTT is used as a screening test for deficiencies of factors VIII, IX and XI. It is recommended that a prolongation to the APTT should be observed when the clotting factor is below 30 IU/dL [1].
- APTT reagents may demonstrate different sensitivities to these clotting factors because of their composition.

AIM

- To compare the APTT sensitivity to clotting factors VIII, IX and XI of two new APTT reagents, **Yumizen G APTT 4** and **Yumizen G APTT Liq 4**, (HORIBA Medical, Montpellier, France), with 6 commonly used APTT reagents.

METHOD

- FVIII deficient plasma (Siemens, Marburg, Germany), FIX and FXI deficient plasma (Precision Biologic, Halifax, Canada) was spiked with normal plasma (CCNRP, Precision Biologic) to obtain plasmas with factor concentrations of 0-90 IU/dL.
- APTT were performed on 5 different days with the 8 APTT reagents.
- Local reference ranges were generated for each reagent using 30 normal donors.
- The intersection of the upper limit of the APTT reference range and concentration of clotting factor determined the reagent sensitivity.
- All testing was performed on Sysmex CS5100i instrumentation (Kobe, Japan).

RESULTS

- Sensitivity to **FVIII** ranged from 31.5- 67.3 IU/dL (Synthasil, Actin FS), median of 43.8 IU/dL.
- Sensitivity to **FIX** ranged from 15.9-36.1 IU/dL (APTT SP, Synthasil), median 27.2 IU/dL. Sensitivity to FIX was greater than 30 IU/dL with Synthasil only.
- Sensitivity to **FXI** from 28.3-55.2 IU/dL (Actin FSL, Synthafax), median 46.1 IU/dL.
- Sensitivity to FVIII and FXI was greater than 30 IU/dL with all reagents, with the exception of FXI using Actin FSL (28.3 IU/dL).
- The sensitivity of **Yumizen G APTT 4** ranged from 26.8- 43.3 IU/dL.
- The sensitivity of **Yumizen G APTT Liq 4** ranged from 29.7- 48.9 IU/dL.
- Inter assay CV was less than 5.5% for all reagents and clotting factors measured.
- Average inter assay CV for **Yumizen G APTT 4** was 3.7%, 3.9%, 2.6% and **Yumizen G APTT Liq4** was 0.9%, 1.3%, 0.9% for FVIII, FIX, FXI respectively.





Figure 1: The sensitivity of APTT4 and APTT Liq4 to FVIII:C (●), FIX:C (■) and FXI (▲) in IU/dL as determined by the intersection between the lower limit of normal (dotted line) and the individual curves. Mean APTT and standard deviation of each point are shown.

| APTT REAGENT | COMPOSITION EA –ellagic acid RB-rabbit brain PL-phospholipid | SENSITIVITY (IU/dL) | | |
|-----------------------------|-----------------------------------------------------------------------|---------------------|------|------|
| | | FVIII | FIX | FXI |
| Yumizen G APTT 4 | Micronised silica RB PL | 40.9 | 26.8 | 43.3 |
| Yumizen G APTT Liq 4 | EA, RB PL | 37.4 | 29.7 | 48.9 |
| Actin FS | EA purified soy phosphatides | 67.3 | 25.5 | 49.3 |
| Actin FSL | EA soy/RB phosphatides | 46.7 | 22.6 | 28.3 |
| APTT SP | Silica synthetic PL | 38.4 | 15.9 | 35.6 |
| STA-PTTA | Silica, RB PL | 47.6 | 29.0 | 35.7 |
| Synthafax | EA synthetic PL | 62.8 | 27.5 | 55.2 |
| Synthasil | Silica synthetic PL | 31.5 | 36.1 | 49.0 |



CONCLUSIONS

- A normal APTT does not always indicate normal haemostasis.
- The sensitivity to FVIII and FXI of most APTT reagents included in this study was considered acceptable however sensitivity to FIX was lower.
- Yumizen G APTT 4** and **Yumizen G APTT Liq 4** results were reproducible and a prolongation to the APTT was observed above 30 IU/dL for FVIII and FXI.
- The source of activator did not appear to affect sensitivity but the phospholipid source, type and concentration influenced the ability to diagnose reduced factor activities, particularly FIX.

REFERENCES

- Clinical Laboratory Standards Institute. One-Stage Prothrombin Time (PT) Test and Activated Partial Thromboplastin Time (APTT) Test; Approved Guideline - Second Edition H47-A2. 2008.

ACKNOWLEDGEMENTS

This study was funded by **HORIBA** Medical

CONTACT INFORMATION

Annette.Bowyer@nhs.net