

Water Distribution Monitor





HORIBA

Explore the future

Automotive Test Systems | Process & Environmental | Medical | Semiconductor | Scientific

HORIBA WATER DISTRIBUTION MONITO Water Distribution Monitor



HORIBA's TW-100 is a process water quality measurement system designed for drinking water and water distribution plants. The TW-100 monitors turbidity, color, free chlorine and pressure. This chemical free method reduces operating costs as it requires minimal maintenance and improves process control.


This compact system is equipped with sensors to monitor four water quality parameters. Optional sensors can be added to monitor as many as seven water quality parameters. No reagents are required for the measurement.

HORIBA WATER DISTR



The instrument is operated with a touch screen panel for easy operation and display of measured values.

Operation panel specifications 320x240 dots Black and white graphic LCD with backlight



The compact size allows for more installation location options and space for operation.

Features:

- Automatic cleaning
- Adjustable measurement frequency
- •Variable sample volume
- Automatic zero calibration
- Internal water leak detection and alarm
- Data storage to compact flash card



Reagent bottle tray for calibration (standard)

[Chart 1]

Standard type

| Parameters | Method | Range | Repeatability | Calibration method |
|----------------|------------------------------|------------------------|------------------|---|
| Turbidity | Transmitted light absorption | 0 to 2 FTU, 0 to 4 FTU | Full scale ±2.5% | Formazin standard solution or polystyrene latex standard solution |
| Color | Transmitted light absorption | 0 to 10 CU, 0 to 20 CU | Full scale ±5.0% | Standard color solution |
| Free chlorine | Polarography | 0 to 2 mg/L | Full scale ±2.5% | DPD colorimetry |
| Water pressure | Diaphragm pressure sensor | 0 to 1 MPa | Full scale ±1.0% | Standard pressure gauge |


Option (each parameter can be added to the standard type)

| Parameters | Method | Range | Repeatability | Calibration method |
|-------------------|-----------------|--------------|------------------|----------------------------------|
| рН | Glass electrode | pH 2 to 12 | ±0.1 pH | pH 7 and pH 9 standard solutions |
| Conductivity | AC bipole | 0 to 50 mS/m | Full scale ±2.0% | KCL standard solution |
| Water temperature | Thermistor | 0 to 50°C | ±0.5°C | Standard thermometer |

System specifications

| Product name | Water Distribution Monitor | | |
|---|---|--|--|
| Model name | TW-100 | | |
| Parameters | Turbidity, color, free chlorine, water pressure (standard); pH, conductivity, water temperature (option) | | |
| Methods | See chart 1. | | |
| Range | See chart 1. | | |
| Repeatability | See chart 1. | | |
| Display | LCD display: 320 x 240 dot B&W graphic LCD with backlight (touch panel type) Standard version simultaneously displays four measurement categories (turbidity, color, free chlorine, and water pressure); Op tional version simultaneously displays up to seven parameters (the four parameters plus pH, conductivity, and water tempera- ture) | | |
| Calibration method | See chart 1. | | |
| Automatic zero calibration | (Turbidity, color, free chlorine) Calibration method: Measurement of filtered sample water Trigger: Internal control (timer) or external control. Interval: 1 to 9999 hours (can be set as desired within that range) Time: 20 to 99 minutes (can be set as desired within that range) | | |
| Automatic cleaning | (Turbidity, color) Cleaning method: Wiper cleaning Trigger: Internal control (timer) or external control. Interval: 5 to 9999 hours (can be set as desired within that range) (Free chlorine) Continuous cleaning with beads grinder | | |
| Sample water conditions | Temperature: 0 to 40°C (must not freeze) Pressure: 0.1 to 0.75 MPa Flow rate: over 100 mL/min Flow rate for measurement: 50 to 100 mL/min * Be sure to flush the sample water before starting the system. * Be sure to install a bypass for the pipe to which the system is attached. * If there is any possibility that the sample water could freeze, please protect the system from cold temperature. | | |
| Ambient temperature/humidity | Temperature; 0 to 40°C; Humidity: 85% or less | | |
| Analog output | Type: Measurement value: Same number as parameters (4 standard, 7 as maximum) Specifications: DC 4 to 20 mA, insulated output (no insulation between parameters), maximum load resistance 600 Ω | | |
| Output signal | Type: Power off, batch alerts, maintenance Details: Power off: Output when power is off. Batch alerts: Cell temperature error, compensation temperature error, cell wiper error, internal communication error, water leak, battery error, upper and lower limit of measurement, light source error, zero calibration, span calibration, analyzer error Maintenance: Output when the system is in maintenance or calibration mode. Specifications: Non-voltage contact output, a contacts Contact ratings: AC 125 V, 0.3 A, DC 30 V, 1 A (depends on resistance load), each output COM is independent. | | |
| Input signal | Type: Cleaning start, zero calibration start, error check sequence start, error water sampling start Details: Cleaning start: Cleaning procedure initiated with closed contact input Zero calibration start: Zero calibration initiated with closed contact input Error discrimination start: Error discrimination initiated with closed contact input Error water sample sampling start: Error water sample sampling initiated with closed contact input Specifications: Non-voltage contact input (open collector contact can be used), insulated input Resistance when ON: Maximum 100 W; Open voltage: DC 26 V; Short-circuit current: Max. 13 mA | | |
| Communications | Interface: RS-232C compliant (*Remote monitoring system is option) | | |
| Functions | Estimation of sample water volume (counting type), internal water leak detection, error water sampling | | |
| Data memory | Measurement values for each parameters are stored in the unit's data memory. Data can be transferred to a CF card. Memory interval: 1 minute or 1 hour Memory schedule: At the top of the hour Data memory longevity: 1 minute intervals: About 10 days; 1 hour intervals: About 1 year. * The latest data is written to up date a series of data. | | |
| Wiring contact | Wiring ground: For power source: 04.5 to 06; For input/output: 012.5 to 014 | | |
| Piping contact | Sample water input: Rc 1/4; Drainage: Rc 1/4; Calibration solution input: Rc 1/8; Error sample sampling aperture: Rc 1/8 | | |
| Construction | Designed for indoor installation. * Housed in a special case (frame) for exterior installation (option). | | |
| Power | AC 100 V to AC 230 V, ±10%, 50/60 Hz | | |
| Power consumption | AC 100 V to AC 120 V: 120 VA; AC 100 V to AC 230 V: 150 VA | | |
| Weight | | | |
| External dimensions | Approx. 15 kg (Measurement unit only) | | |
| | 350 mm (W) x 160 mm (D) x 420 mm (H) (not including installation fixtures and projecting portions) | | |
| Paint color/Material Installation environment | Munsell 5PB8/1; Unit case is made of SUS. Install on a flat surface without vibration shocks. No dust, mist, or corrosive gases in the atmosphere. Lower air pressure. No direct sunlight. Ventilated air. •2,000 m or less above sea level. | | |

Exterior dimensions (unit: mm)



The HORIBA Group adopts IMS (Integrated Management System) which integrates Quality Management System ISO9001, Environmental Management System ISO14001, and Occupational Health and Safety Management System OHSAS18001. We have now integrated Business Continuity Management System ISO22301 in order to provide our products and services in a stable manner, even in emergencies

Please read the operation manual before using this product to assure safe and proper handling of the product.

•The specifications, appearance or other aspects of products in this catalog are subject to change without notice.

•Please contact us with enquiries concerning further details on the products in this catalog.

China

Thailand

Vietnam

Indonesia

•The color of the actual products may differ from the color pictured in this catalog due to printing limitations.

It is strictly forbidden to copy the content of this catalog in part or in full.

•The screen displays shown on products in this catalog have been inserted into the photographs through compositing.

•All brand names, product names and service names in this catalog are trademarks or registered trademarks of their respective companies.

HORIBA KOREA Ltd.

| HOR | BA |
|-------------|----|
| Head Office | |

2 Miyanohigashi, Kisshoin Minami-ku, Kyoto, Japan Phone: 81 (75) 313-8123 Fax: 81 (75) 321-5725 http://www.horiba.com e-mail: info@horiba.co.jp

HORIBA (China) Trading Co., Ltd.

Unit D, 1F, Building A, Synnex International Park, 1068 West Tianshan Road, Shanghai, 200335, China Phone: 86 (21) 6289-6060 Fax: 86 (21) 6289-5553 Beijing Office

12F, Metropolis Tower, No.2, Haidian Dong 3 Street, Beijing, 100080, China

Phone: 86 (10) 8567-9966 Fax: 86 (10) 8567-9066

HORIBA (Thailand) Limited

East Office 850 / 7 Soi Lat Krabang 30 / 5, Lat Krabang Road, Lat Krabang, Bangkok 10520, Thailand Phone: 66 (0) 2734 4434 Fax: 66 (0) 2734 4438

HORIBA Instruments (Singapore) Pte Ltd. Singapore

3 Changi Business Park Vista #01-01 Akzonobel House, Singapore 486051 Phone: 65 (6) 745-8300 Fax: 65 (6) 745-8155

HORIBA Vietnam Co., Ltd.

Unit 6, 10 Floor, CMC Tower, Duy Tan Street, Dich Vong Hau Ward, Cau Giay District, Hanoi, Vietnam Phone: 84 (24) 3795-8552 Fax: 84 (24) 3795-8553

PT HORIBA Indonesia

Bulletin:HAE-T0250A

Jl. Jalur Sutera Blok 20A, No.16-17, Kel. Kunciran, Kec. Pinang Tangerang-15144, Indonesia Phone: 62 (21) 3044-8525 Fax: 62 (21) 3044-8521

Manufactured by HORIBA Advanced Techno

Seoul Branch 10, Dogok-Ro, 6-Gil, Gangnam-Gu, Seoul-Si, 06259, Korea Phone: 82 (2) 753-7911 Fax: 82 (2) 756-4972 **HORIBA India Private Limited** India 246, Okhla Industrial Estate, Phase 3 New Delhi-110020, India Phone: 91 (11) 4646-5000 Fax: 91 (11) 4646-5020 Technical Center D-255, Chakan MIDC Phase-II, Bhamboli Village, Pune-410501, India Phone: 91 (21) 3567-6000 Bangalore Office No.55, 12th Main, Behind BDA Complex, 6th sector, HSR Layout, Bangalore South, Bangalore-560102, India Phone: 91 (80) 4127-3637 **HORIBA Instruments Incorporated** 9755 Research Drive, Irvine, CA 92618, U.S.A Phone: 1 (949) 250-4811 Fax: 1 (949) 250-0924

Houston Office 5390 Bay Oaks Drive, Pasadena, TX 77505 Phone: 1 (281) 482- 4334 Fax: 1 (281) 674-6058

HORIBA Instruments Brazil, Ltda.

Rua: Presbitero Plinio Alves de Souza, 645, Loteamento Polo Multivias Barirro Medeiros-Jundiai Sao Paulo CEP 13.212-181 Brazil Phone: 55 (11) 2923-5400 Fax: 55 (11) 2923-5490

HORIBA FRANCE SAS

Korea

USA

Brazil

Les Ulis Office 12. Av des Tropiques Hightec Sud, F-91955 Les Ulis, France Phone: 33 (1) 69-29-96-23 Fax: 33 (1) 69-29-95-77

HORIBA UK Limited

Northampton Office Kyoto Close Moulton Park, Northampton NN3 6FL, UK Phone: 44 (1604) 542-500 Fax: 44 (1604) 542-699

HORIBA Europe GmbH

Germany Hans-Mess-Str.6 D-61440 Oberursel German Phone: 49 (6172) 1396-0 Fax: 49 (6172) 1373-85 Leichlingen Office Julius-kronenberg Str.9 D-42799 Leichlingen Germany Phone: 49 (2175) 8978-0 Fax: 49 (2175) 8978-50

HORIBA Czech

Prague Office Prumyslova 1306 / 7, CZ-10200, Praha 10, Czech Republic Phone: 420 (2) 460-392-65

HORIBA (Austria) GmbH

Kaplanstrasse 5 A-3430 Tulln, Austria Phone: 43 (2272) 65225 Fax: 43 (2272) 65230 Austria

Czech

France

UK

Printed in Japan 1707SK62

HORIBA